



TRADE AND COMMERCE  
CANADA

STANDARDS DIVISION

OTTAWA, December 5, 1956.

TYPE APPROVAL

SPRAGUE TYPE NO.3  
POSITIVE DISPLACEMENT GAS METER

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type No.3 Positive Displacement Gas Meter, manufactured by the Sprague Meter Company, Bridgeport, Connecticut, U.S.A.

Rating of Apparatus:

Badged Capacity .....	320 cu.ft./hr.(air)
Differential Pressure at badged capacity ....	0.5" water gauge
Capacity per revolution .....	0.294 cu.ft.
Connections (male) .....	1½"

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: The instrument approved herewith is essentially the same in construction, design and appearance as the Sprague meters previously approved for use in Canada.

Note: Approval covers the aluminumcase as well as the ironcase meter; also the use of front and back covers made of cold rolled steel, when required.

*E. F. Power*

E. F. Power,  
Assistant Director (E&G),  
Standards Division.

*R. W. MacLean*  
R. W. MacLean,  
Director,  
Standards Division.

Ref: A-597