



TRADE AND COMMERCE
CANADA

SD-GA.117

STANDARDS DIVISION

OTTAWA, February 27, 1957.

TYPE APPROVAL

CANADIAN METER COMPANY "30M" TINCASE
POSITIVE DISPLACEMENT GAS METER

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: The "30M" Tincase Positive Displacement Gas Meter, manufactured and distributed by the Canadian Meter Company Limited, Milton, Ontario.

Rating of Apparatus:

Badged Capacity	720 cu.ft./hr.(air)
Differential Pressure at badged capacity	0.5" water gauge
Capacity per revolution625 cu.ft.
Connections	1 1/4" male.

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: Instrument identical to type approved in Circular SD-GA.10, March 28, 1949. (Re-submitted for "air" rating).

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

Ref: A-628