



TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

OTTAWA, July 10, 1957.

TYPE APPROVAL

ROCKWELL NO. 1600
POSITIVE DISPLACEMENT GAS METER

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Rockwell No. 1600 Positive Displacement Gas Meter, manufactured by the Rockwell Manufacturing Company, Pittsburgh, Pa., and distributed in Canada by its subsidiary, the Rockwell Manufacturing Company of Canada Ltd., Guelph, Ontario.

Rating of Apparatus:

Badged Capacity	620 cu.ft./hr.(air)
Differential Pressure at badged capacity	0.5" water gauge
Capacity per revolution	0.385 cu.ft.
Connections (male)	1 1/4" or 1 1/2"

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: This meter is similar in all respects to the Rockwell type 800 previously approved on Circular SD-GA.87 of October 14, 1955 and on Circular SD-GA.109 of November 30, 1956. The only change is in the "type" designation.

E. F. Power
E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

Ref: A-450