



TRADE AND COMMERCE  
CANADA

STANDARDS DIVISION

OTTAWA, July 15, 1957.

TYPE APPROVAL

ROCKWELL NO. 5000 ALUMINUMCASE  
POSITIVE DISPLACEMENT GAS METER

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Rockwell No. 5000 Positive Displacement Gas Meter, manufactured by the Rockwell Manufacturing Company, Pittsburgh, Pa., and distributed in Canada by its subsidiary, the Rockwell Manufacturing Company of Canada Ltd., Guelph, Ontario.

Rating of Apparatus:

- Badged Capacity ..... 1740 cu.ft./hr.(air)
- Differential Pressure at badged capacity ..... 0.5" water gauge
- Capacity per revolution ..... 2 cu.ft.
- Connections (female) ..... 4"

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: This meter is similar in all respects to the Rockwell No. 4 meter previously approved on Circular SD-GA.112 of January 17, 1957. The only change is in the "type" designation.

*E. F. Power*

E. F. Power,  
Assistant Director (E&G),  
Standards Division.

*R. W. MacLean*

R. W. MacLean,  
Director,  
Standards Division.

Ref: A-616