



TRADE AND COMMERCE
CANADA

SD.GA.130

STANDARDS DIVISION

OTTAWA, August 13, 1957.

TYPE APPROVAL

PARKINSON & COWAN TYPE "IN4"
POSITIVE DISPLACEMENT GAS METER

The apparatus specified and illustrated herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "IN4" Positive Displacement Gas Meter, manufactured by Parkinson & Cowan (Gas Meters) Limited, London, England, and distributed in Canada by Duragas Equipment Limited, 300 Greenwood Avenue, Toronto 6, Ontario.

Rating of Apparatus:

Badged Capacity 2700 cu.ft./ hr.(air)
Differential Pressure at Badged Capacity ... 0.5" water gauge
Capacity per revolution 6.25 cu.ft.
Connections 4" flange

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: This type of meter is dry lubricated and no external oiling is necessary. The bearings are made from bronze, graphite-impregnated. The meter case is constructed of heavy gauge tinned plate. The valves are made of Phenolic-graphite and the diaphragms of leather.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

Ref: A-644