

TRADE AND COMMERCE  
CANADA

SD.GA.131

## STANDARDS DIVISION

OTTAWA, August 13, 1957.

TYPE APPROVALROCKWELL NO. 4½ PRESSED STEEL CASE  
POSITIVE DISPLACEMENT GAS METER

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Rockwell No. 4½ Positive Displacement Gas meter, manufactured by the Rockwell Manufacturing Company, Pittsburg, Pa., and distributed in Canada by its subsidiary, the Rockwell Manufacturing Company of Canada Ltd., Guelph, Ontario.

## Rating of Apparatus:

Badged Capacity .....	2800 cu.ft./hr.(air)
Differential Pressure at badged capacity ...	0.5" water gauge
Capacity per revolution .....	3.33 cu.ft.
Connections (female) .....	4"

Applications: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: This meter is similar in all respects to the Emco 4½ previously approved on Circular SD.GA.3 of October 20, 1948, the only change being a re-rating on an "air basis".

E. F. Power,  
Assistant Director (E&G),  
Standards Division.

R. W. MacLean,  
Director,  
Standards Division.

Ref: A-621