



TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

SD-GA.134

OTTAWA, October 29, 1957.

TYPE APPROVAL

BEGWACO BANTAM "1200" STEELCASE
POSITIVE DISPLACEMENT GAS METER

The apparatus specified and illustrated herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Begwaco Bantam "1200" Steelcase Positive Displacement Gas Meter, manufactured and distributed in Canada by Begwaco Meters (Canada) Limited, No. 40 Highway, Chatham, Ontario.

Rating of Apparatus:

Badged Capacity 950 cu.ft./hr.(air)
Differential Pressure at badged capacity 0.5" water gauge
Capacity per revolution 0.5 cu.ft.
Connections 1½" female.

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: Welded steel case, with die cast diaphragm chambers and gas way. Bakelite valves and seats. Leather or synthetic diaphragms.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

Ref: A-704