



TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

OTTAWA May 6, 1958.

TYPE APPROVAL

CANADIAN METER COMPANY TYPE "AL-110"
POSITIVE DISPLACEMENT GAS METER

The apparatus specified and illustrated herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "AL-110" Low Pressure Positive Displacement Gas Meter, manufactured and distributed by the Canadian Meter Company Limited, Milton, Ontario.

Rating of Apparatus:

Badged Capacity	150 cu.ft./hr.(air)
Differential Pressure at badged capacity	0.5" water gauge
Capacity per revolution	0.111 cu.ft.
Connections	female side connections $\frac{1}{2}$ " or $\frac{3}{4}$ " F.P.T.

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: Except for the flat top cover and female side connections, this meter is basically identical to the "A-225" approved on June 12, 1957, under Circular SD-GA.122.

E. F. Power
(for) E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

Ref: A-752