



TRADE AND COMMERCE
CANADA

SD-GA.141

STANDARDS DIVISION

OTTAWA July 3, 1958.

TYPE APPROVAL

AMERICAN METER COMPANY "BP 1200" BASE PRESSURE INDEX

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Metric Base Pressure Index "BP 1200", manufactured by the American Meter Company Limited, Erie, Pa., U.S.A., and distributed in Canada by the Canadian Meter Company Limited, Milton, Ontario.

Rating of Apparatus:
Maximum Gauge Pressure 1200 p.s.i.

Description: This device is identical to the instrument approved in N.R.C. Circular #1, May 26, 1933, except that the pressure range has been increased to 1200 p.s.i. This approval shall be considered as an extension to Circular SD-GA.39, July 6, 1953.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. Maclean
R. W. Maclean,
Director,
Standards Division.

Ref: A-768A