



TRADE AND COMMERCE
CANADA

SD-GA.143

STANDARDS DIVISION

OTTAWA..... July 10, 1958.....

MODIFICATION OF TYPE APPROVAL

SMITH TYPE "U.D.Mc.1"
POSITIVE DISPLACEMENT GAS METER

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Type "U.D.Mc.1" Die-Cast Aluminium Positive Displacement Gas Meter, manufactured by Smith Meters Ltd., London, England, and distributed in Canada by Smith Meters (Canada) Limited, Toronto 18, Ontario.

Rating of Apparatus:

Budget Capacity	250 cu.ft./hr. (air)
Differential Pressure at budget capacity	0.5" water gauge
Capacity per revolution	0.3 cubic foot
Connections (male)	1 1/2"

Modification: The modification consists only in a structural change in the top casting, the inlet and outlet connections now being 6" centres. The meter was approved in Circular SD-GA.103, September 10, 1956, for the above budget capacity.

E. F. Power

E. F. Power,
Assistant Director (BS&G),
Standards Division.

R. W. Maclean
R. W. Maclean,
Director,
Standards Division.

Ref: A-339G

✓