

TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

OTTAWA, November 25, 1958.

MODIFICATION OF TYPE APPROVALROCKWELL TYPES "150", "210", "250", "310" AND "415"
POSITIVE DISPLACEMENT GAS METERS

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Rockwell Types "150", "210", "250", "310" and "415"
Positive Displacement Gas Meters, manufactured by the Rockwell
Manufacturing Company of Canada Ltd., Guelph, Ontario.

Modification: The above-noted types of meters were originally fitted with a wire screen soldered in the inlet port to prevent access to the working parts. The modification consists of substituting a single bar guard, permanently fixed in the inlet port, for the previously-approved wire screen.

E. F. Power
(E. F. Power),
Assistant Director (E&G),
Standards Division.

R. W. MacLean
(R. W. MacLean),
Director,
Standards Division.

Ref: A-792