



Consumer and  
Corporate Affairs Canada

Consommation  
et Corporations Canada

Legal Metrology  
Branch

Direction de la  
Métrologie légale

Holland Avenue  
Ottawa, Ontario  
K1A 0C9

Avenue Holland  
Ottawa, Ontario  
K1A 0C9

Your file    *Votre référence*

Our file    *Notre référence*

Telephone: (613)952-0639

File N°: 06635  
Project: AP-GL-91-0018

Mr. Greg Milenchuk  
North Western Utilities Ltd.  
10035 - 105th Street  
Edmonton, Alberta  
T5J 2V6

1991-07-19

Dear Sir,

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notices of Approval SD-GA.152 (80B, 250B, and 500B meters), S-GA.251 (AL-800 meter), G-29 (AL1000 meter), G-126 (AL1400 meter), G-127 (AL2300 meter) and G-128 (AL5000 meter). It is noted that these devices are to be modified as follows:

- For a meter found with a broken shaft, which holds the tangent T-C element to the tangent mounting bracket, this shaft may be replaced with an 8-32 threaded screw, after appropriately drilling and tapping the bracket.

I am advised by the Gas Metrology Laboratory that their evaluation of the proposed modifications indicates that they may be regarded as immaterial in relation to the approval provisions of the Act. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

W.R. Virtue, Chief  
Legal Metrology Laboratories

cc: Approvals Mailing List

Canada