



TRADE AND COMMERCE  
CANADA

SD-GA.158

## STANDARDS DIVISION

OTTAWA..... June 16, 1959.....

MODIFICATION OF TYPE APPROVALROCKWELL TYPE #210 POSITIVE DISPLACEMENT GAS METER

The modifications to the apparatus specified herein have been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Type #210 Positive Displacement Gas Meter, manufactured and distributed in Canada by the Rockwell Manufacturing Company of Canada Limited, Guelph, Ontario.

## Rating of Apparatus:

Badged Capacity ..... 160 cu.ft./hr.(air)  
Differential Pressure at badged capacity ..... 0.5" water gauge  
Capacity per revolution ..... .0625 cu.ft.  
Connections ..... 1 $\frac{1}{4}$ " and 1" male.

## Modifications:

1. Cover - The cover design has been modified so that the index box cover will be interchangeable with that used on the Rockwell types #150, L.P.G., #250, #310 and #415 meters.
2. Body - The body design has been altered so that it will be similar in appearance to the Rockwell types #250, #310 and #415 meters.
3. Sealing - The sealing arrangement will be similar to that of the Rockwell types #150, L.P.G., #250, #310 and #415 meters.

The type #210 meter was previously approved on Circulars SD-GA.105 of October 1, 1956 and SD-GA.108 of November 29, 1956.

E.F. Power,  
Assistant Director (E&G),  
Standards Division.

R. W. MacLean,  
Director,  
Standards Division.

Ref: A-792B