

TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

OTTAWA, July 9, 1959.

MODIFICATION OF TYPE APPROVALROCKWELL #5 POSITIVE DISPLACEMENT GAS METER

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Rockwell #5 Positive Displacement Gas Meter, manufactured by the Rockwell Manufacturing Company, Pittsburgh, Pa., and distributed in Canada by its subsidiary, the Rockwell Manufacturing Company of Canada, Ltd., Guelph, Ontario.

Modification: The intermediate gear ratio has been changed to permit the use of the type of index normally fitted to meters #750, #2 $\frac{1}{2}$, #3, #4 and #4 $\frac{1}{2}$, which have a ten-cubic-foot-per-revolution test dial. The minimum test run volume of fifty cubic feet corresponds to nine tangent revolutions.

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean,
Director,
Standards Division.

Ref: A-621B