



TRADE AND COMMERCE  
CANADA

STANDARDS BRANCH

OTTAWA November 18, 1959.

TYPE APPROVAL

CANADIAN METER COMPANY TYPE "AL-110"  
POSITIVE DISPLACEMENT GAS METER

The apparatus specified herein has been duly approved by the Standards Branch under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "AL-110" Low-Pressure Positive Displacement Gas Meter, manufactured and distributed by the Canadian Meter Company Limited, Milton, Ontario.

Rating of Apparatus:

Badged Capacity .....	150 cu.ft./hr.(air)
Differential Pressure at badged capacity .....	0.5" water gauge
Capacity per revolution .....	0.111 cu.ft.
Connections .....	$\frac{1}{2}$ " or $\frac{3}{4}$ " F.P.T.

Application: Measurement in distribution service of L.P.G. gases only.

Description: The type "AL-110 meter was previously approved under SD-GA.138 of May 6, 1958 and incorporated the use of a 4-figure clock index having a maximum reading of 9999 thousand cubic feet. The meter is manufactured specially for the L.P.G. industry and the manufacturer has been authorized to use the cyclometer index approved under SD-GA.40 of July 31, 1953. This latter index has a maximum reading of 99990 cubic feet. Either index may be used when required.

*E. F. Power*

E. F. Power,  
Assistant Director (E&G),  
Standards Branch.

*R. W. MacLean*  
R. W. MacLean,  
Director,  
Standards Branch.

Ref: A-752