



TRADE AND COMMERCE
CANADA

8-GA.237

STANDARDS BRANCH

OTTAWA August 16, 1962.

TYPE APPROVAL

CANADIAN METER COMPANY TYPE "AL-800"
ALUMINUMCASE POSITIVE DISPLACEMENT GAS METER

The apparatus specified herein has been duly approved by the Standards Branch under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "AL-800" Aluminumcase Positive Displacement Gas Meter, manufactured and distributed in Canada by the Canadian Meter Company Limited, Milton, Ontario.

Rating of Apparatus:

Badged Capacity	650 cu.ft./hr. (air)
Differential Pressure at badged capacity ...	0.5" water gauge
Capacity per revolution	0.385 cu.ft.
Working Pressure	100 p.s.i.
Connections (female)	1½"

Application: Measurement in distribution service of manufactured, natural and petroleum gases or mixtures thereof.

Description: The meter approved herein is similar to the meter approved in Circular SD-GA.156, May 26, 1959. The case is constructed to withstand the increased working pressure and the connections are 1½" female.

E. F. Power

E. F. Power,
Chief, Electricity and Gas Division,
Standards Branch.

R. W. MacLean
R. W. MacLean,
Director,
Standards Branch.

Ref: A-815A