

TRADE AND COMMERCE
CANADA

STANDARDS BRANCH

S-GA.239

OTTAWA, August 31, 1962.

TYPE APPROVALCANADIAN METER COMPANY TYPE "AL-425"
POSITIVE DISPLACEMENT GAS METER

The apparatus specified herein has been duly approved by the Standards Branch under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "AL-425" Aluminum Case Positive Displacement Gas Meter, manufactured and distributed in Canada by the Canadian Meter Company Limited, Milton, Ontario.

Rating of Apparatus:

Badged Capacity	340 cu.ft./hr.(air)
Working Pressure	10 p.s.i.
Differential Pressure at badged capacity	0.5" water gauge
Capacity per revolution	0.200 cu.ft.
Connections (male)	1", 1 $\frac{1}{4}$ " and 1 $\frac{1}{2}$ "

Application: Measurement in distribution service of heating gases.

Description: The meter approved herein is identical to the type "AL-425" approved in Circulars S-GA.190 with 1 $\frac{1}{4}$ " male connections and S-GA.199 with 1 $\frac{1}{2}$ " male connections, except that the meter may be fitted with 1" male connections.

E. F. Power

E. F. Power,
Chief, Electricity and Gas Division,
Standards Branch.

R. W. MacLean
R. W. MacLean,
Director,
Standards Branch.

Ref: A-868