

Note: The illustration on the back of this circular shows the type 415 meter with incomplete temperature compensation information. The following is required:-

- (i) the index dial label shall read "Temp. Comp. 60°F Base".
- (ii) the nameplate shall include inscription "Temp. Comp. to 60°F Base".

R. W. MacLean
(for) R. W. MacLean,
Director,
Standards Branch.

G. E. Anderson
G. E. Anderson,
Chief Engineer,
Standards Branch.

Ref. SL-100-765 E.