

TRADE AND COMMERCE
CANADA

STANDARDS BRANCH

S-GA.172

OTTAWA, November 12, 1959.

MODIFICATION OF TYPE APPROVALROCKWELL TYPE #415 POSITIVE DISPLACEMENT GAS METER

The modification to the apparatus specified and illustrated herein has been duly approved by the Standards Branch under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Rockwell Tyre #415 Positive Displacement Gas Meter, manufactured and distributed in Canada by Rockwell Manufacturing Company of Canada Limited, Guelph, Ontario.

Modification: Formerly the type #415 gas meter was sealed by passing the sealing wire through two heads of the hand-hole cover screws and two heads of the index box cover screws. This modification permits a better arrangement whereby the sealing wire passes through one cover-to-body cap screw or bolt at the front of the meter in addition to the screws previously included in the sealing. Henceforth the new arrangement shall be used.

The type #415 meter is covered by Approval Circulars SD-GA.59 of December 20, 1954 and SD-GA.164 of July 8, 1959.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Branch.

R. W. MacLean
R. W. MacLean,
Director,
Standards Branch.

Ref: A-379C

ROCKWELL TYPE #415 POSITIVE DISPLACEMENT GAS METER



