



TRADE AND COMMERCE  
CANADA

STANDARDS BRANCH

S-GA-178

*(Replaced by  
S-GA-178 amended  
Feb 26, 1962  
last rev.)*

OTTAWA January 27, 1960.

TYPE APPROVAL

ROOTS-CONNERSVILLE TYPES "3M25" AND "7M25"  
ROTARY POSITIVE DISPLACEMENT GAS METERS

The apparatus specified and illustrated herein has been duly approved by the Standards Branch under the provisions of the Gas Inspection Act, Chap. 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Types "3M25" and "7M25" Rotary Positive Displacement Gas Meters, manufactured by Roots-Connorsville Blower Division of Dresser Industries Limited, Connorsville, Indiana, U.S.A. and distributed in Canada by Roots-Connorsville Blower (Canada) Limited, 629 Adelaide St. West, Toronto, Ontario.

Rating of Apparatus:	"3M25"	"7M25"
Badged Capacity (cu. ft. per hr. / air) . . . . .	3000	7000
Capacity per revolution (cu. ft.) . . . . .	0.02205	0.06120
Working Pressure (p.s.i.) . . . . .	125	125
Connections . . . . .	Flanged	Flanged

Description: The meters are all cast iron with "O" ring joint construction. The one (1) pair of timing gears is steel with steel bearing and brass planetary reduction gears on the opposite end. The counter is chain-driven with a stainless steel chain from the planetary gear shaft. Both gear ends of the meter are splash-lubricated, and bullseye-type oil sight gauges are provided on each end. The meter is designed to mount in a vertical position; if installed in a horizontal service line, the counter will be upside down.

Note: The type "3M25" meter may be tested in a meter shop, if requested, using the bell-type prover, in which case the regular fee provided in Part I of the "Schedule of Inspection Fees for the Testing of Gas and Gas Meters" shall apply. When tested 'in situ', the fee provided for Rotary Type Positive Displacement Meters shall apply.

*E. F. Power*

E. F. Power,  
Assistant Director (E&G),  
Standards Branch.

*R. W. MacLean*  
R. W. MacLean,  
Director,  
Standards Branch.

Ref: A-856



ROOTS-COMNERSVILLE ROTARY POSITIVE DISPLACEMENT GAS METER



