



TRADE AND COMMERCE  
CANADA

STANDARDS BRANCH

OTTAWA January 13, 1961.

TYPE APPROVAL

PARKINSON COWAN (CANADA) LTD. CLASS B250  
POSITIVE DISPLACEMENT GAS METER

The apparatus specified herein has been duly approved by the Standards Branch under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Parkinson Cowan (Canada) Ltd. Class B250 Positive Displacement Gas Meter, manufactured by Parkinson Cowan (Canada) Ltd., 830 Richmond Street, Chatham, Ontario.

Rating of Apparatus:

Badged Capacity ..... 200 cu.ft./hr.(air)  
Differential Pressure at badged capacity ..... 0.5" water gauge  
Capacity per revolution ..... 0.11 cu.ft.  
Connections ..... 1" male.

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: The meter approved herein is identical to the Begwaco "Bantam" Class B250 positive displacement gas meter previously manufactured by Begwaco Meters (Canada) Limited. The meter is now manufactured by Parkinson Cowan (Canada) Ltd. Except for the name, the meter is unchanged.

*E. F. Power*

E. F. Power,  
Chief, Electricity & Gas Division,  
Standards Branch.

*R. W. MacLean*  
R. W. MacLean,  
Director,  
Standards Branch.

Ref: A-799

*Aluminum index cover  
changed to transparent Plexiglas  
Jan 23/63.*

