



TRADE AND COMMERCE
CANADA

STANDARDS BRANCH

OTTAWA March 21, 1963.

TYPE APPROVAL

NEGRETTI AND ZAMBRA "MERSTEEL" TEMPERATURE RECORDER

The apparatus specified herein has been duly approved by the Standards Branch under the provisions of the Gas Inspection Act, Chapter 129, R. S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: The "Mersteel" Temperature Recorder, manufactured and distributed in Canada by Negretti and Zambra (Canada) Limited, Toronto 14, Ontario.

Application: Measurement in distribution service of manufactured, natural and petroleum gases, or mixtures thereof, where temperature is a factor in the billing.

Rating of Apparatus:
Temperature Range Any range between -50°F and +150°F
Temperature Span not greater than 160°F
Bulb Diameter 1/4" to 3/4"
Capillary Tubing Length 20 feet maximum

Description: This single pen recorder is similar in all respects to the type approved under Circular S-GA.204, dated February 17, 1961 with the exception that the temperature range is extended.

E. F. Power
E. F. Power,
Chief, Electricity and Gas Division,
Standards Branch.

R. W. MacLean
R. W. MacLean,
Director,
Standards Branch.

Ref: A-895B

