



Date: July 23, 2004

File number: 26922-AP-AM-04-0022

**MODIFICATION ACCEPTANCE LETTER (MAL)  
FOR  
Approved Weighing and Measuring Devices**

**Type of Device**

Automatic and Non-Automatic Electronic Prepacking Scale

**Purpose**

The purpose of this letter is to convey details of modifications to weighing and measuring devices which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered within the scope of Notice of Approval AM-5293 and any subsequent revisions to this Notice of Approval.

**Applicant**

Bizerba Canada Inc.  
2810 Argentia Road  
Mississauga, Ontario  
L5N 8L2

**Description of Modifications**

- These devices are also available in avoirdupois unit of measure (lb).

* Max <sub>1</sub> : 0 - 6 lb Max <sub>2</sub> : 6 - 15 lb	e <sub>1</sub> : 0.002 lb e <sub>2</sub> : 0.005 lb	* Max <sub>1</sub> : 0 - 15 lb Max <sub>2</sub> : 15 - 30 lb	e <sub>1</sub> : 0.005 lb e <sub>2</sub> : 0.01 lb
* Max <sub>1</sub> : 0 - 6 lb Max <sub>2</sub> : 6 - 15 lb Max <sub>3</sub> : 15 - 30 lb	e <sub>1</sub> : 0.002 lb e <sub>2</sub> : 0.005 lb e <sub>3</sub> : 0.01 lb	Max : 6 lb Max : 15 lb Max : 30 lb	e : 0.002 lb e : 0.005 lb e : 0.01 lb

\* Multi-interval device

Original signed by:

Michel Maranda  
Technical Coordinator  
Mass Measurement Discipline



Date: 23 juillet 2004

Dossier: 26922-AP-AM-04-0022

**LETTRE D'ACCEPTATION DE MODIFICATION (LAM)**  
**pour**  
**Appareils de pesage et de mesure approuvés**

**Type d'Appareil**

Balance de pré-emballage électronique automatique et non-automatique

**Objet**

La présente lettre vise à expliquer les détails des modifications apportées aux appareils de pesage et de mesure dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée de l'avis d'approbation AM-5293 et de toutes révisions subséquentes.

**Requérant**

Bizerba Canada Inc.  
2810 Argentia Road  
Mississauga, Ontario  
L5N 8L2

**Descriptions des modifications**

- Ces appareils sont également disponible en unité de mesure avoirdupois (lb).

* Max <sub>1</sub> : 0 - 6 lb Max <sub>2</sub> : 6 - 15 lb	e <sub>1</sub> : 0,002 lb e <sub>2</sub> : 0,005 lb	* Max <sub>1</sub> : 0 - 15 lb Max <sub>2</sub> : 15 - 30 lb	e <sub>1</sub> : 0,005 lb e <sub>2</sub> : 0,01 lb
* Max <sub>1</sub> : 0 - 6 lb Max <sub>2</sub> : 6 - 15 lb Max <sub>3</sub> : 15 - 30 lb	e <sub>1</sub> : 0,002 lb e <sub>2</sub> : 0,005 lb e <sub>3</sub> : 0,01 lb	Max : 6 lb Max : 15 lb Max : 30 lb	e : 0,002 lb e : 0,005 lb e : 0,01 lb

\* Appareil à échelons multiples

Copie authentique signée par :

Michel Maranda  
Coordonnateur en technologie  
Discipline des masses