



Date: February 14, 2007

File number: 26686-AP-AM-06-0116

**MODIFICATION ACCEPTANCE LETTER (MAL)
FOR
Approved Weighing and Measuring Devices**

Type of Device

Electronic Bench/ Platform Scale

Purpose

The purpose of this letter is to convey details of modifications to weighing and measuring devices which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered within the scope of Notice of Approval AM-5404 and any subsequent revisions to this Notice of Approval.

Applicant

A&D Engineering Inc.
1555 Mc Candless Drive,
Milpitas, California,
USA 95035

Description of Modifications

The purpose of this modification is to correct the device's subtype from multi-interval to multiple range. The table on the following page reflects the correction and replaces the table in the Notice Of Approval on page 3 of 7.

Model/Modèle	Platform/Plate-forme	Max ⁱ and/et e ^j		
HV-15KGV HV-15KGL HV-15KV-WP	250 mm × 250 mm	0 to/à 3 kg × 0.001 kg 0 kg to/à 6 kg × 0.002 kg 0 kg to/à 15 kg × 0.005 kg	0 to/à 6 lb × 0.002 lb 0 lb to/à 15 lb × 0.005 lb 0 lb to/à 30 lb × 0.01 lb	0 to/à 96 oz × 0.05 oz 0 oz to/à 240 oz × 0.1 oz 0 oz to/à 480 oz × 0.2 oz
HV-60KGV HV-60KGL HV-60KV-WP	330 mm × 424 mm	0 to/à 15 kg × 0.005 kg 0 kg to/à 30 kg × 0.01 kg 0 kg to/à 60 kg × 0.02 kg	0 to/à 30 lb × 0.01 lb 0 lb to/à 60 lb × 0.02 lb 0 lb to/à 150 lb × 0.05 lb	0 to/à 480 oz × 0.2 oz 0 oz to/à 960 oz × 0.5 oz 0 oz to/à 2400 oz × 1 oz
HV-200KGV HV-100KGL HV-200KV-WP	390 mm × 530 mm	0 to/à 60 kg × 0.02 kg 0 kg to/à 150 kg × 0.05 kg 0 kg to/à 220 kg × 0.1 kg	0 to/à 150 lb × 0.05 lb 0 lb to/à 300 lb × 0.1 lb 0 lb to/à 500 lb × 0.2 lb	0 to/à 2400 oz × 1 oz 0 oz to/à 4800 oz × 2 oz 0 oz to/à 8000 oz × 5 oz

Original signed by:
Michel Maranda
Technical Coordinator
Mass Measurement Discipline



Date: 14 février, 2007

Dossier: 26686-AP-AM-06-0116

**LETTRE D'ACCEPTATION DE MODIFICATION (LAM)
pour
Appareils de pesage et de mesure approuvés**

Type d'Appareil

Balance de table/ plate-forme électronique

Objet

La présente lettre vise à expliquer les détails des modifications apportées aux appareils de pesage et de mesure dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée de l'avis d'approbation AM-5404 et de toutes révisions subséquentes.

Requérant

A&D Engineering Inc.
1555 Mc Candless Drive,
Milpitas, Californie,
É. U. 95035

Descriptions des modifications

1. Cette modification vise à corriger le sous-type de l'appareil d'échelons multiples à étendues multiples. Cette modification est illustrée dans le tableau suivant qui remplace celui de la page 3 de 7 dans l'avis d'approbation.

Model/Modèle	Platform/Plate-forme	Max ¹ and/et é		
HV-15KGV HV-15KGL HV-15KV-WP	250 mm × 250 mm	0 to/à 3 kg × 0.001 kg 0 kg to/à 6 kg × 0.002 kg 0 kg to/à 15 kg × 0.005 kg	0 to/à 6 lb × 0.002 lb 0 lb to/à 15 lb × 0.005 lb 0 lb to/à 30 lb × 0.01 lb	0 to/à 96 oz × 0.05 oz 0 oz to/à 240 oz × 0.1 oz 0 oz to/à 480 oz × 0.2 oz
HV-60KGV HV-60KGL HV-60KV-WP	330 mm × 424 mm	0 to/à 15 kg × 0.005 kg 0 kg to/à 30 kg × 0.01 kg 0 kg to/à 60 kg × 0.02 kg	0 to/à 30 lb × 0.01 lb 0 lb to/à 60 lb × 0.02 lb 0 lb to/à 150 lb × 0.05 lb	0 to/à 480 oz × 0.2 oz 0 oz to/à 960 oz × 0.5 oz 0 oz to/à 2400 oz × 1 oz
HV-200KGV HV-100KGL HV-200KV-WP	390 mm × 530 mm	0 to/à 60 kg × 0.02 kg 0 kg to/à 150 kg × 0.05 kg 0 kg to/à 220 kg × 0.1 kg	0 to/à 150 lb × 0.05 lb 0 lb to/à 300 lb × 0.1 lb 0 lb to/à 500 lb × 0.2 lb	0 to/à 2400 oz × 1 oz 0 oz to/à 4800 oz × 2 oz 0 oz to/à 8000 oz × 5 oz

Copie authentique signée par :
Michel Maranda
Coordonnateur en technologie
Discipline des masses