



Date: January 17, 2005

File number: 26686-AP-AM-04-0109

**MODIFICATION ACCEPTANCE LETTER (MAL)
FOR
Approved Weighing and Measuring Devices**

Type of Device

Electronic Bench/Platform Scale

Purpose

The purpose of this letter is to convey details of modifications to weighing and measuring devices which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered within the scope of Notice of Approval AM-5477 and any subsequent revisions to this Notice of Approval.

Applicant

Mettler-Toledo
1150 Dearborn Drive
Worthington, Ohio, 43085

Description of Modifications

- Add models WHN 060P and WHN 150P and update platter dimensions for models WHN 060L and WHN 150B

Models	Max	e_{min}	Load cell capacity	Platter dimensions
WHN 060L	60 kg / 150 lb 50 kg / 100 lb	0.02 kg / 0.05 lb 0.01 kg / 0.02 lb	100 kg / 220 lb	457 mm by 457 mm
WHN 060P	60 kg / 150 lb 50 kg / 100 lb	0.02 kg / 0.05 lb 0.01 kg / 0.02 lb	100 kg / 220 lb	457 mm by 610 mm
WHN 150B	150 kg / 300 lb 100 kg / 250 lb	0.05 kg / 0.1 lb 0.02 kg / 0.05 lb	200 kg / 440 lb	457 mm by 457 mm
WHN 150P	150 kg / 300 lb 100 kg / 250 lb	0.05 kg / 0.1 lb 0.02 kg / 0.05 lb	200 kg / 440 lb	457 mm by 610 mm

Original signed by:
Michel Maranda
Technical Coordinator
Mass Measurement Discipline



Date: 17 janvier 2005
Dossier: 26686-AP-AM-04-0109

LETTRE D'ACCEPTATION DE MODIFICATION (LAM)
pour
Appareils de pesage et de mesure approuvés

Type d'Appareil

Balance électronique de table / à plate-forme

Objet

La présente lettre vise à expliquer les détails des modifications apportées aux appareils de pesage et de mesure dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée de l'avis d'approbation AM-5477 et de toutes révisions subséquentes.

Requérant

Mettler-Toledo
1150 Dearborn Drive
Worthington, Ohio, 43085

Descriptions des modifications

- Ajout des modèles WHN 060P et WHN 150P et actualisation des dimensions du plateau des modèles WHN 060L et WHN 150B

Modèles	Max	e_{\min}	Capacité de cellule de pesage	Dimensions de plateau
WHN 060L	60 kg / 150 lb 50 kg / 100 lb	0.02 kg / 0.05 lb 0.01 kg / 0.02 lb	100 kg / 220 lb	457 mm sur 457 mm
WHN 060P	60 kg / 150 lb 50 kg / 100 lb	0.02 kg / 0.05 lb 0.01 kg / 0.02 lb	100 kg / 220 lb	457 mm sur 610 mm
WHN 150B	150 kg / 300 lb 100 kg / 250 lb	0.05 kg / 0.1 lb 0.02 kg / 0.05 lb	200 kg / 440 lb	457 mm sur 457 mm
WHN 150P	150 kg / 300 lb 100 kg / 250 lb	0.05 kg / 0.1 lb 0.02 kg / 0.05 lb	200 kg / 440 lb	457 mm sur 610 mm

Copie authentique signée par :
Michel Maranda
Coordonnateur en technologie
Discipline des masses