



Date: June 10, 2008

File number: 26686-APAM080026

**MODIFICATION ACCEPTANCE LETTER (MAL)
FOR
Approved Weighing and Measuring Devices**

Type of Device

Electronic Weighing and Load Receiving Element

Purpose

The purpose of this letter is to convey details of modifications to weighing and measuring devices which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered within the scope of Notice of Approval **AM-5558** and any subsequent revisions to this Notice of Approval.

Applicant

Bizerba Canada Inc.
2810 Argentia Road, #9
Mississauga, Ontario
L5N 8L2

Description of Modifications

- The purpose of this modification is to add the equivalent lb units to the Devices main metrological characteristics table in the Notice Of Approval.

Model Modèle	Max	ϵ_{min}	Model Modèle	Max	ϵ_{min}	Model Modèle	Max	ϵ_{min}	Model Modèle	Max	ϵ_{min}
20 VE-S	6 kg (15 lb)	0.001 kg (0.002 lb)	50 VE-S	0 - 6 kg	0.002 kg	100 VE-S	30 kg (60 lb)	0.01 kg (0.02 lb)	200 VE-S	150 kg (300 lb)	0.05 kg (0.1 lb)
				6 - 15 kg (0-15 lb) (15-30 lb)	0.005 kg (0.005 lb) (0.01 lb)		60 kg (150 lb)	0.02 kg (0.05 lb)			
							0 - 30 kg 30 - 60 kg (0 - 60 lb) (60 - 150 lb)	0.005 kg 0.01 kg (0.02 lb) (0.05 lb)			

Original signed by:

Jean Lemay
Technical Coordinator
Mass Measurement Discipline



Date: 10 juin 2008

Dossier: 26686-APAM080026

LETTRE D'ACCEPTATION DE MODIFICATION (LAM)
pour
Appareils de pesage et de mesure approuvés

Type d'Appareil

Dispositif peseur et récepteur de charge électronique

Objet

La présente lettre vise à expliquer les détails des modifications apportées aux appareils de pesage et de mesure dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée de l'avis d'approbation **AM-5558** et de toutes révisions subséquentes.

Requérant

Bizerba Canada Inc.
2810 Argentia Road, #9
Mississauga, Ontario
L5N 8L2

Descriptions des modifications

- Cette modification vise à ajouter l'équivalent en unité lb dans le tableau des caractéristiques métrologiques principales des appareils de l'avis d'approbation.

Model Modèle	Max	e _{min}	Model Modèle	Max	e _{min}	Model Modèle	Max	e _{min}	Model Modèle	Max	e _{min}
20 VE-S	6 kg (15 lb)	0.001 kg (0.002 lb)	50 VE-S	0 - 6 kg 6 - 15 kg (0-15 lb) (15-30 lb)	0.002 kg 0.005 kg (0.005 lb) (0.01 lb)	100 VE-S	30 kg (60 lb)	0.01 kg (0.02 lb)	200 VE-S	150 kg (300 lb)	0.05 kg (0.1 lb)
							60 kg (150 lb)	0.02 kg (0.05 lb)			
							0 - 30 kg 30 - 60 kg (0 - 60 lb) (60 - 150 lb)	0.005 kg 0.01 kg (0.02 lb) (0.05 lb)			

Copie authentique signée par:
Jean Lemay
Coordonnateur en technologie
Discipline des masses