



STANDARDS DIVISION

OTTAWA, January 7, 1949

NOTIFICATION OF TYPE LISTING

Under the provisions of the appropriate Act the apparatus specified and illustrated herein has been listed as a Class A device, and may be used in Canada in accordance with the conditions applicable to this classification.

Authority: Weights and Measures Act, Chapter 212, R.S. 1927, and Regulations thereunder (P.C. 3459).

Apparatus Listed: "Peerless" Portable Platform Beam Type Scale, manufactured by The Scale Shop, Toronto, Ontario.

Rating of Apparatus: 500 lb. or 1000 lb. maximum capacity (2 models).


Application: Weighing of commodities in general trade.

Conditions: As prescribed in P.C. 3459.

Description: This is a standard type of floor-platform-beam scale operated by a lever system. It is available in either single beam or double beam models and the beams are brass, un-notched, with aluminum insert. A sliding poise with a set-screw locating device is provided.

Testing: The standard tests applicable to a platform, beam scale will apply.


J. D. Staver,
Chairman, Examining Board.


A. F. Gill,
Director of Standards.

"PEERLESS" PORTABLE PLATFORM BEAM TYPE SCALE

THE PEERLESS SCALE COMPANY
ALBANY, N. Y.

