

TRADE AND COMMERCE
CANADA

SD-WA.36

STANDARDS DIVISION

OTTAWA, January 14, 1949

NOTIFICATION OF TYPE LISTING

Under the provisions of the appropriate Act, the apparatus specified and illustrated herein has been listed as a Class C device, and may be used in Canada in accordance with the conditions applicable to this classification.

Authority: Weights and Measures Act, Chapter 212, R.S. 1927 and Regulations thereunder (P.C. 3459).


Apparatus Listed: Model 500 Electrograph Scale, manufactured by the Exact Weight Scale Company, Columbus, Ohio.


Rating of Apparatus: 0.5 gram minimum to 15 gram maximum capacity.

Conditions: As prescribed in Part II, Sections 3 and 4 of P.C. 3459.

Description: This scale is equipped with an even balanced lever mounted on agate bearings; an electrically dampened lever connected to an electric amplifying circuit; hardened steel knife edges; a potentiometer for altering the balancing point; a manually operated tare control for adjusting tare weight; a 4 in. x 4 in. platter; and a non-linear type of dial showing the exact weight point. It is designed for fine weighings on a production basis; i.e., packaging in drug manufacturing plants, etc.


Testing: Not applicable, as Class C devices do not require government inspection.


J. L. Stiver,
Chairman, Examining Board.


A. F. Gill,
Director of Standards.

EXACT WEIGHT ELECTROGRAPH SCALE

MODEL 500



FRONT VIEW

