

TRADE AND COMMERCE
CANADA

SD-WA.58

STANDARDS DIVISION

PLEASE QUOTE FILE.....

OTTAWA, May 8th, 1950.....

NOTIFICATION OF TYPE LISTING

Under the provisions of the appropriate Act, the apparatus specified and illustrated herein has been listed as a Class A device, and may be used in Canada in accordance with the conditions stated herein.

Authority: Weights and Measures Act, Chapter 212, R.S. 1927, and Regulations thereunder (P.C. 3459).

Apparatus Listed: Model 213-NB Over and Under Scale, manufactured by the Exact Weight Scale Company, Columbus, Ohio.


Rating of Apparatus: 3 lbs.

Application: Weighing of commodities for prepackaging purposes.

Conditions: As prescribed in P.C. 3459 with the added mandatory requirement that this scale shall not be used on the retail premises of any business enterprise.

Description: The design and mechanism of this scale is identical with that of the Model 213 listed under SD-WA.28 of January 7th, 1949, excepting that the weighing capacity is 2 lbs. less, the indicator tower has been raised 4 inches and the pull-rod lengthened accordingly; two beams are provided - one calibrated to 8 oz. x $\frac{1}{2}$ oz. and the other calibrated to $\frac{1}{2}$ oz. x 1/100 oz.; two adjustable pointers are furnished on the chart and are set manually - one to indicate the desired net weight and the other to indicate the desired tolerance; between these pointers, the weight indicating pointer should come to rest.

Testing: The standard tests for an equal-arm, beam type scale will apply and notice should be taken of the requirements of Part VI, Section 28, of P.C. 3459 respecting prepackaging tolerances.


R. W. MacLean,
Director,
Standards Division.

EXACT WEIGHT PREPACKAGING SCALE

MODEL 213-NB

