

Application:

Mail Weighing and the determination of postal charges.

Description:

This device is a simple lever beam type scale with the platform being maintained in a horizontal position by a checking device. The beam is notched and the beam poise is fitted with a dog for engaging in the notches. The steel beam is of hollow construction and the balance ball is operated by a screw located within the beam and operated from the beam tip by a screwdriver. An over and under knife edge indicator is inserted in the beam near the beam tip. Pivots and bearings are hardened steel and are self-aligning. The platform, scale casing, and base are manufactured of pressed sheet steel. Models listed are of similar design.

Testing:


The standard tests for postal scales shall apply.

Reference: G6922-T167-20

Conditions of Approval:

Weights and Measures Act, Chapter 212, R.S.C. 1927, and Regulations thereunder (P.C. 3459). As prescribed in P.C. 3459, and to be stamped or engraved with the words "Postal Scale" as laid down in the Weights and Measures Act, Section 69.

Under the provisions of the appropriate Act, the apparatus specified and illustrated herein has been listed as a Class A device, and may be used in Canada in accordance with the conditions stated herein.

  
-----  
John Armstrong  
Chief  
Weights and Measures Division  
Legal Metrology Branch