



CANADA

SD-WA.85

DEPARTMENT OF TRADE AND COMMERCE  
STANDARDS DIVISION

IN YOUR REPLY REFER TO FILE NO.....

OTTAWA, April 13th, 1951.....

NOTIFICATION OF LISTING

Under the provisions of the appropriate Act, the apparatus specified and illustrated herein has been listed as a Class A device, and may be used in Canada in accordance with the conditions applicable to this classification.

Authority: Weights and Measures Act, Chapter 212, R.S. 1927, and Regulations thereunder (P.C. 3459).

Apparatus Listed: Counter computing cylindrical chart scale Model 4515, equipped with chart GM, GO, GR or GT and Model 4505, equipped with chart GN, GP or GS, manufactured by the Toledo Scale Company of Canada, Limited, Windsor, Ontario.


Rating of Apparatus: Both models may have capacities of 18 lbs., 24 lbs. or 30 lbs.


Application: Weighing commodities in retail trade.

Conditions: As prescribed in P.C. 3459.

Description: This device is a counter computing cylinder chart scale of the spring and lever type. The scale is provided with two levers operating at right angles to one another. Resistance to load is provided by 2 pairs of temperature compensated springs. Pivots and bearings are of hardened steel and are self aligning. Movement is transmitted to the chart cylinder by rack and pinion. Zero adjustment is located in the scale column behind the flashing on the platter. Model 4505 is a straight read and Model 4515 is an angle read scale.

Testing: The standard tests for a counter computing cylinder chart type scale shall apply.

  
G. S. Phillips,  
Assistant Director (W&M),  
Standards Division.

  
R. W. MacLean,  
Director,  
Standards Division.

Reference A-183

TOLEDO SCALE COMPANY  
"Defiance" Model 4515

