

TRADE AND COMMERCE
CANADA

SD-WA.119

STANDARDS DIVISION

OTTAWA, August 27th, 1952.

NOTIFICATION OF TYPE LISTING

Under the provisions of the Weights and Measures Act, Chapter 36, S.C. 1951 and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus: Pitney-Bowes Parcel Post Scale, Model S-250; an automatic indicating scale of the fan chart type, manufactured by the National Store Specialty Co., Bareville, Pa., U.S.A.

Rating of Apparatus: 30 lbs. capacity.

Application: Weighing of parcels for mailing.

Conditions: As prescribed in P.C. 6894. This scale only shows computations and red pound figures to 25 lbs. Weights up to 30 lbs. may, however, be weighed. Consequently this scale is rated at 30 lbs. capacity.

Description: This device is a platform scale of the fan chart type. It is equipped with a single main lever, hardened steel pivots and agate bearings. The load receiving member is maintained in position by a stabilizing link which extends up into the chart housing pillar and is connected to a cross member, which is anchored to the housing.

The chart indicator registers to 15 lbs. when the side beam poise is in the zero notch, and will register up to 30 lbs. when this poise is in the second notch. Chart values over 15 lbs. are printed in red figures, smaller in size than those used for values below 15 lbs.

Testing: The standard tests for an automatic indicating scale of the fan chart type will apply.

C. S. Phillips,
Assistant Director (W&M),
Standards Division.

R. W. MacLean,
Director,
Standards Division.

Reference: A-222

PITNEY-BOWERS PARCEL POST SCALE

Model B-250

