



TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

OTTAWA, December 7, 1955.

APPROVAL LISTING

Bench Type Platform Scales - Manufactured by
W. & T. Avery Limited, Birmingham, England.

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein have been listed as approved devices and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Platform bench type beam scale, Models 3302AAA and 3302AAE.

Rating of Apparatus: Capacities of 100 lbs., 250 lbs. and 300 lbs.

Application: Weighing commodities in general trade.

Conditions: As prescribed in P.C. 6894.

Description: These devices are platform scales with a single beam graduated to 5 lbs. with 1 ounce graduations. The lever system is suspended from links and consists of two A type levers, one lever having an extension arm for transmitting power to the beam rod. The pivots are V type with round shanks set in reamed holes in the levers and are adjustable. Bearings are self aligning hardened pressed steel inserts. The base, deck, pillar, bracket and trig loop are cast iron. Model 3302AAE differs from 3302AAA in that the pillar is longer.

Testing: The standard tests for a platform scale shall apply.

C.S. Phillips,
Assistant Director (W&M),
Standards Division.

R.W. MacLean,
Director,
Standards Division.

Reference: A-361

SD-WA.241

AVERY BENCH SCALE

Model 3302AAA

