



CANADA

SD-WA. 262

DEPARTMENT OF TRADE AND COMMERCE

STANDARDS DIVISION

OTTAWA, March 20th, 1956.

APPROVAL LISTING

Bench type platform scales Manufactured by
W & T Avery, Birmingham, England.

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein have been listed as approved devices and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Platform bench type beam scale, Models 3302 ABC, and 3302 ABP and 3302 ABW with capacities of 120 lbs, 300 lbs. and 300 lbs. respectively.

Application: Weighing commodities in general trade.

Conditions: As prescribed in P.C. 6894.

Description: These devices are platform scales equipped with full capacity compound beam. One beam bar is graduated to 10 lbs. x 1 oz. The second beam bar which is notched may be graduated in 10 lb. increments. The reading face of the beam is in the horizontal plane and the leverage system is the same as that used in Model 3302 AAA (SD - WA. 241). Model 3302 ABC and 3302 ABP differ in capacity only and 3302 ABW is equipped with a long beam supporting column.

Testing: The standard tests for a platform scale shall apply.

C. S. Phillips,
Assistant Director (W&M),
Standards Division.

R. W. MacLean,
Director,
Standards Division.

Reference: A 384

AVERY PLATFORM SCALE
MODEL 5302 ABC

