



TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

OTTAWA...February 6, 1959.....

APPROVAL LISTING

FAIRBANKS - SINGLE ANIMAL SCALE

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Platform scale with full capacity beam, Model 6700A, manufactured by Canadian Fairbanks-Morse Company, Limited, Sherbrooke, P.Q.

Rating of Apparatus: 3000 lbs. capacity.


Application: General farm use and single animal weighing when equipped with a cattle rack on the live deck.


Conditions: As prescribed in P.C. 6894.

NOTE: It is required that the scale be located in a fixed position on a suitable level concrete slab or foundation, or other suitable permanent and level surface.

Description: This device is a platform scale equipped with a full capacity beam. The main levers are contained within a welded frame composed of ten inch channels welded at the four corners. The under deck leverage system consists of two pipe levers with bar steel side extensions connected to the connecting rod which in turn is connected to a shelf lever. Load and fulcrum main bearings are of the closed stirrup type. The beam is a full capacity beam with a notched section of 3000 lbs. x 100 lbs. The capacity of the beam butt fractional section is 100 lbs. x 1 lb. graduations. The balance ball has an approximate range of 300 lbs.

Testing: The standard tests for a platform scale shall apply.


C.S. Phillips,
Assistant Director (W&M),
Standards Division.


R.W. MacLean,
Director,
Standards Division.

Reference: A-514.

FAIRBANKS - MORSE
SINGLE ANIMAL AND FARM USE SCALE

SD-WA. 362

MODEL 6700A

