

TRADE AND COMMERCE
CANADA

SD-WA. 365

STANDARDS DIVISION

OTTAWA, February 18, 1959.

APPROVAL LISTING

SANITARY-BIZERBA, COMPUTING COUNTER SCALE

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Computing Counter Scale Model OP-15 with a capacity of 15 lbs., manufactured by Sanitary Scale Company, Belvidere, Illinois, U.S.A.

Application: Weighing commodities in general trade.


Conditions: As prescribed in P.C. 6894


Description: This device is a Computing Counter Scale with optical type registration and is similar in basic design to Model OP-15-P listed on SD-WA. 364 except that:

- (1) The taring knob has been removed and replaced by a slot in the spindle to allow zero adjustment by a screwdriver. The range of adjustment has also been limited.
- (2) Weight values are shown on the customer's side of the scale.

The price range is .10 cents to \$1.59 per pound and is set manually.

Testing: The standard tests for a counter computing scale shall apply.


C.S. Phillips,
Assistant Director (W&M),
Standards Division.


R.W. MacLean,
Director,
Standards Division.

SANITARY-BIZBBA
COMPUTING COUNTER SCALE
MODEL OP-15

SD-WA. 365

