



TRADE AND COMMERCE
CANADA

SD-WA. 369

STANDARDS DIVISION

OTTAWA, February 23, 1959.

APPROVAL LISTING

OHAUS - FOUR IN ONE GRAIN TEST SCALE

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Four in One Grain tester Model 14, manufactured by Ohaus Scale Corporation, 1050 Commerce Avenue, Union, New Jersey, U.S.A.

Rating of Apparatus: Maximum capacity 550 grams.

Application: Testing of grain - weight per bushel, establishing dockage and direct weighing.


Conditions: As prescribed in P.C. 6894 with the added requirement that the measuring bucket when supplied with the scale be marked with the same serial number as the scale.


Description: This device is a redesigned version of the model 14 grain tester listed on SD-WA. 97.

Changes incorporated in the scale are as follows:

- (1) The base has been redesigned to give it a modern appearance.
- (2) The poises and markings on the beams have been changed to provide better readability and facilitate weighing.
- (3) The zero ball adjustment assembly has been relocated to provide better protection.
- (4) The beam butt has been reinforced.

Testing: The standard tests for a grain tester of this type shall apply.


C.S. Phillips,
Assistant Director (W&M),
Standards Division.


R.W. MacLean,
Director,
Standards Division.

Reference: A-191F

CHAUS 4 in 1 GRAIN TESTING SCALE

SD-MA. 369

MODEL 14

