



SD-WA. 376

TRADE AND COMMERCE  
CANADA

## STANDARDS DIVISION

OTTAWA, May 25, 1959.

## APPROVAL LISTING

## NATIONAL - OVER AND UNDER PLATFORM COUNTING SCALES

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein have been listed as approved devices and may be used in Canada in accordance with the conditions applicable.

**Apparatus Listed:** Bench and floor type platform scales for weighing and counting; Models 205 and 205T each 150-pounds capacity, Models 206, 206T, 206 NA and 206 NAT each 1,000-pounds capacity, manufactured by National Scale Company, 43 Mills Street, Springfield, Mass. U.S.A.

**Application:** Weighing commodities in general trade and counting in industry and warehouses.

**Conditions:** As prescribed in P.C. 6894.


**Description:** These devices are platform and beam type counting and weighing scales with over and under indicator. The 205 models are bench type and the 206 models are portable floor type. The suffix letters NA indicate that registering head and beams are aligned with the length center line of the base of the scale. The suffix T denotes counting ratios of 999:1, 99:1, 9:1, 1000:1, 100:1, and 10:1 otherwise the ratios are 99:1, 9:1, 100:1, 10:1, 144:1, 12:1. The underdeck leverage system is similar to that of Model 97 (SD-WA. 177).

**Testing:** The standard tests for a platform scale shall apply and in addition the counting accuracy and sensitivity should be determined by checking known loads on the deck with appropriate weights placed in the counterbalance scoops.

*C.S. Phillips*  
C.S. Phillips,  
Assistant Director (W&M),  
Standards Division.

*R.W. MacLean*  
R.W. MacLean,  
Director,  
Standards Division.


Reference: A-530 and A-269



NATIONAL


CORRECTING SCALE

MODEL 205



OVER AND UNDER AND

CORRECTING HEAD ASSEMBLY



SD-11, 376

1.1

C

C

