



Date: July 29, 2009

File number: AP-AV-09-0017

**MODIFICATION ACCEPTANCE LETTER (MAL)  
FOR  
Approved Measuring Devices**

**Type of Device**

Register

**Purpose**

The purpose of this letter is to convey details of modifications to liquid measuring devices which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

**Scope**

The modifications described in this letter affect devices identified in the following Notice(s) of Approval:

**Approval Number**

AV-2342

**Applicant**

Liquid Controls  
A unit of IDEX Corporation  
105 Albrecht Drive  
Lake Bluff, Illinois, 60044-2242

**Description of Modifications**

- To add software version SR217.

Original copy signed by:

Amina Abid  
Manager-Liquids Measurement



Date: 29 juillet, 2009

Dossier: AP-AV-09-0017

**LETTRE D'ACCEPTATION DE MODIFICATION (LAM)**  
**pour**  
**Compteurs et appareils approuvés**

**Type d'appareil**

Enregistreur

**Objet**

La présente lettre vise à expliquer les détails des modifications apportées aux compteurs dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée des avis d'approbation d'origine énumérés dans le présent document et de toutes leurs révisions subséquentes.

**Portée**

Les modifications décrites dans la présente lettre s'appliquent aux appareils visés par l'avis d'approbation suivant:

**Numéro d'approbation**

AV-2342

**Requérant**

Liquid Controls  
A unit of IDEX Corporation  
105 Albrecht Drive  
Lake Bluff, Illinois, 60044-2242

**Description des modifications**

- Pour ajouter de version de logiciel SR217.

Copie authentique signée par :

Amina Abid  
Gestionnaire de mesure des liquides