



Measurement Canada
An agency of Industry Canada

Mesures Canada
Un organisme d'Industrie Canada

MAL No. - NE De LAM

MAL-V23

Date: July 31, 2000

File number: AP-AV-00-0005

MODIFICATION ACCEPTANCE LETTER (MAL) FOR Approved Measuring Devices

Type of Device

Electronic Register/Control System for Bulk Meters

Purpose

The purpose of this letter is to convey details of modifications to liquid measuring devices which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

Scope

The modifications described in this letter affect devices identified in the following Notice(s) of Approval:

Approval Number

AV-2327, AV-2304T , AV-2229

Applicant

Rosemount Instruments Ltd.
6280 Danville Road.
Mississauga, Ontario
L5T 2H7

Description of Modifications

The models approved under the above noted Notices of Approval can be used in conjunction with any approved and compatible coriolis type mass flow meter.

Original copy signed by:

Randy Byrtus
Approvals Technical Coordinator
Fluid Measurement Discipline



Measurement Canada

An agency of Industry Canada

Mesures Canada

Un organisme d'Industrie Canada

MAL No. - NE De LAM

MAL-V23

Date : 31 juillet, 2000

Dossier : AP-AV-00-0005

LETTRE D'ACCEPTATION DE MODIFICATION (LAM)

pour

Compteurs et appareils approuvés

Type d'appareil

Enregistreur électronique/Dispositif de régulation

Objet

La présente lettre vise à expliquer les détails des modifications apportées aux compteurs dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée de l'avis d'approbation d'origine énuméré dans le présent document et de toutes ses révisions subséquentes.

Portée

Les modifications décrites dans la présente lettre s'appliquent aux compteurs visés par l'avis d'approbation suivant :

Numéro d'approbation

AV-2327, AV-2304T , AV-2229

Requérant

Rosemount Instruments Ltd.
6280 Danville Road.
Mississauga, Ontario
L5T 2H7

Description des modifications

Les modèles approuvés en vertu des avis d'approbation indiqués ci-dessus peuvent être utilisés de concert avec n'importe quel débitmètre massique approuvé et compatible de type Coriolis.

Copie authentique signée par:

Randy Byrtus
Coordonnateur technique d'approbations
Discipline de la mesure des fluides