



Measurement Canada
An agency of Industry Canada

Mesures Canada
Un organisme d'Industrie Canada

MAL No. - NE De LAM
MAL-V26

Date: September 21, 2000

File number: AP-AV-00-0018

MODIFICATION ACCEPTANCE LETTER (MAL) FOR Approved Measuring Devices

Type of Device

Electronic Register/Control System

Purpose

The purpose of this letter is to convey details of modifications to liquid measuring devices which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

Scope

The modifications described in this letter affect devices identified in the following Notice(s) of Approval:

Approval Number

AV-2304T

Applicant

Rosemount Instruments Ltd.
407 West Vine Street
Hatfield, PA
USA 19440

Description of Modifications

The devices approved under the above mentioned Notice of Approval can be equipped with firmware version DE1. This version is identical to the previously approved version, "DD", except that it permits automatic temperature compensation for liquefied petroleum gas having a density of 580 kg/m³ at 15°C.

Original copy signed by:

Randy Byrtus
Approvals Technical Coordinator
Fluid Measurement Discipline



Measurement Canada
An agency of Industry Canada

Mesures Canada
Un organisme d'Industrie Canada

MAL No. - NE De LAM
MAL-V26

Date: 21, septembre 2000

Dossier: AP-AV-00-0018

LETTRE D'ACCEPTATION DE MODIFICATION (LAM) **pour** **Compteurs et appareils approuvés**

Type d'appareil

Enregistreur électronique/Dispositif de régulation

Objet

La présente lettre vise à expliquer les détails des modifications apportées aux compteurs dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée des avis d'approbation d'origine énumérés dans le présent document et de toutes leurs révisions subséquentes.

Portée

Les modifications décrites dans la présente lettre s'appliquent aux compteurs visés par l'avis d'approbation suivant:

Numéro d'approbation

AV-2304T

Requérant

Rosemount Instruments Ltd.
407 West Vine Street
Hatfield, PA
USA 19440

Description des modifications

Les appareils approuvés en vertu de l'avis d'approbation susmentionné peuvent être équipés d'un microprogramme, version DE1. Cette version est identique à la version antérieure approuvée « DD » en plus de permettre la compensation de température automatique des gaz de pétrole liquéfiés ayant une masse volumique de 580 kg/m³ à 15°C.

Copie authentique signée par:

Randy Byrtus
Coordonnateur en technologie, approbations
Mesure des fluides