



TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

OTTAWA, June 17, 1958.

APPROVAL LISTING

ROCKWELL ROTOCYCLE BULK METER

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Positive displacement rotary type bulk meter, Models 400, 600 and 1000 with maximum rated capacities of 400 G.P.M., 600 G.P.M. and 1000 G.P.M. respectively, manufactured by Rockwell Manufacturing Company., Pittsburg, Pennsylvania, U.S.A. and Guelph, Ontario.


Application: Metering bulk quantities of liquid in trade.


Conditions: As prescribed in P.C. 6894.

Description: These devices are rotary type volumetric meters identical to Models 4 BLX, 5 BLX and 6 BLX listed on SD-WA.286 except that the gear type linkage used to synchronize the movement of the buckets with the rotor movement is replaced by a crank type linkage.

NOTE: The aforementioned BLX meters may be submitted for verification when equipped with the crank type synchronizing linkage.

Testing: The standard tests for a bulk meter shall apply.


G. S. Phillips,
Assistant Director (W&M)
Standards Division.


R. W. MacLean,
Director,
Standards Division.

Reference: A-408A

MODEL 600

