



CANADA

SD-WA.452

DEPARTMENT OF TRADE AND COMMERCE
STANDARDS BRANCH

OTTAWA, March 14, 1963.

APPROVAL LISTING

GILBERT AND BARKER - GASOLINE DISPENSER

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Remote pumping dispenser with "T" shaped casing, Model MD 20 and may also be used with remote hose stands models HC 2 with internal retracting hose and HC2-1 with outside retracting hose.

Manufactured by Gilbert and Barker Manufacturing Company Limited, 64 Jefferson Avenue, Toronto, Ontario.

Rating of Apparatus: 12 G. P. M.


Application: Retail gasoline trade.


Conditions: As prescribed in P.C. 6894.

Description: Component parts similar to Model 44RC (SD-WA.401). Model MD 20 is "T" shape with the meter directly below the counter mechanism. This model is used with remote hose stands. The hose stand bears the visi-glass and the hoses are identified by number and the same number is clearly shown on the register just below the dial, e.g. " This dial for No. 1 hose".

Note:- A limited number of installations were authorized using pumps, model 44DRC with the remote hose stands HC 2 and HC2-1. However, future installations of this type will use model MD 20 dispenser.

Testing: The standard tests for a retail gasoline dispenser will apply.


C. S. Phillips,
Chief, Weights & Measures Division,
Standards Branch.


R. W. MacLean,
Director,
Standards Branch.

Reference: A-553

GILBERT AND BARKER - GASOLINE DISPENSER

SD-V-4,152

Hose Stand
(Model HC 2)

MODEL NO 20

