



Consumer and
Corporate Affairs

Consommation et
corporations

Standards

Normes

RECORD OF APPROVAL ACTION

TEMPORARY Approval Listing Number - S.WA- T-114

(Issued in advance of preparation of draft Approval Listing only in case of Interim Approval).

Ottawa, March 7, 1979

Company Requesting Approval - Lockheed Electronics Company
Industrial Technology Division
U.S. Route 22, Plainfield, N.J. 07061


<u>Model</u>	<u>Description</u>
LCR-E	This register indicates volume on a mechanical roller-counter as on previous models, but the price setting and computation is performed electronically.

Conditions: This temporary approval is limited to one installation on truck No. 14269 at Thermoshell Division, 3975 Keele St., Toronto, incorporating a 2-inch Neptune Red Seal Meter.

There is to be a non-return leakproof check valve installed to prevent the dilated hose from squeezing liquid back through the meter. A pressure gauge is to be installed between the meter and the hose to monitor the sealing of the check valve.

The truck shall be re-routed as necessary away from residential areas into industrial areas where the electronic computer is more likely to be subjected to radio frequency interference from such sources as arc welders, high frequency welders, induction heaters, RF transmitters, truck-mounted C B radios, etc. A ticket shall be printed at each delivery and a copy of each ticket shall be forwarded to the Chief, Weights & Measures, Ottawa.

This temporary approval is valid until May 30, 1979 except that it may be revoked at any prior time if the subject device fails to perform satisfactorily or any of the conditions prescribed or any requirements of the Regulations are not met. R.N.G. Equipment Ltd., 32 Staffell Drive Rexdale, shall be held responsible for ensuring that the conditions of this temporary approval are complied with.


John Armstrong
Chief
Weights and Measures Division
Legal Metrology Branch

File No.

Note: File Record of Approval Action with Approval Listing notices when Approval Listing number is assigned in advance of preparation of notice; file at back of book by Central File No. when Approval Listing number is not assigned.

Remove Record of Approval Action from Approval Listing Book when the company is notified of rejection, or when Special Approval Listing notice or draft of Approval Listing notice is filed with notices; replace draft of Approval Listing notice with printed notice when available.



Consumer and
Corporate Affairs Canada

Consommation et
Corporation Canada

Consumer Standards
Directorate

Direction générale
des normes

RECORD OF APPROVAL ACTION

Approval Listing Number - S.WA-T114 Addendum #4
(Temporary)

(Issued in advance of preparation of formal Approval Notice)

Ottawa, April 8, 1980

Company Requesting Approval - Lockheed Electronics Company
Industrial Technology Division
U.S. Route 22, Plainfield, N.J. 07061

Type of Device -


<u>Model</u>	<u>Description</u>
LCR-E	This register indicates volume with a mechanical roller-counter as on previous models, but the price setting and computation is done electronically.

Conditions: This temporary approval is extended to one installation on a Gulf Canada vehicle tank at Toronto, Ontario.

There is to be a non-return leakproof check valve installed to prevent the dilated hose from squeezing liquid backwards through the meter. A pressure gauge is to be installed between the check valve and the hose to monitor the sealing of the hose.

A ticket shall be printed at each delivery and a copy of the ticket forwarded to the Chief of Weights and Measures, Ottawa. This temporary approval is valid until June 30, 1980 under the same conditions noted in the original issue of S.WA-T114.

File No. 6953-L271


.....
J.P. Buchanan
Acting Chief
Weights & Measures Division
Legal Metrology Branch