



Consumer and
Corporate Affairs Canada

Consommation et
Corporation Canada

Ottawa/Hull
K1A 0C9

Consumer Standards
Directorate

Direction générale
des normes

RECORD OF APPROVAL ACTION

Temporary Approval listing Number - SWA-T133 Rev.1

(Issued in advance of preparation of draft Approval Listing only
in case of interim Approval).

Ottawa, December 21, 1979

Company Requesting Approval - Brooks Instruments Canada Limited
P.O. Box 150, Markham, Ontario

Type of Device - Automatic Temperature Compensator, model 4400
for use with Brooks Bi-Rotor meters on liquids
for which ATC has been authorized.


Application for Approval - The temporary approval applies to
Brooks meters with model 4400 ATCs as follows:

1. Sunoco Refinery, Sarnia, on a 3-inch meter for LPG.
2. Gibson Petroleum, Edmonton, on two 4" meters for LPG.

Conditions of Approval - All installations shall comply with the
(draft) document "Additional Specifications for Liquid Meters
Equipped with Automatic Temperature Compensators," with
particular reference to SVM 8.4.1 to 8.4.9. The ATC is to be
checked once a week during the period of temporary approval by
measuring the product temperature with a certified thermometer
and comparing actual and theoretical NET readings with GROSS
readings. All data are to be recorded and reported monthly to
the Chief, Weights and Measures.

This temporary approval expires on October 9, 1980.

File No. 6953-B365-12


.....
John Armstrong
Chief
Weights & Measures Division
Legal Metrology Branch