

TRADE AND COMMERCE
CANADA

S-WA.403

STANDARDS BRANCH

Ottawa, January 13, 1961

APPROVAL LISTING

TOKHEIM - BULK METER

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.O. 6894), the apparatus specified and illustrated herein have been listed as approved devices and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Volumetric 1 1/2" bulk meter and ancillary equipment. Models 632, 632TP, 634, 634TP and 635TP manufactured by Tokheim Corporation, Fort Wayne, Indiana, U.S.A.

Note: Model numbers vary due to type of additional equipment supplied with basic meter such as: built in filter and air eliminator, ticket printer, non printing register, preset mechanism and built in check back pressure valve.

Rating of Apparatus: 32 G.P.M. maximum.


Application: Tank truck meter and general trade use.


Conditions: As prescribed in P.O. 6894.

Description: This device is a positive displacement meter with three pistons operating vertically in separate cylinders. The connecting rods of the pistons are connected to a wobble type plate which controls (1) the sequence of operation of the pistons (2) through a converting mechanism operates the counter (3) regulates the operation of the valves.

The metering unit is encased in a two part casing.

Testing: The standard tests applicable to a bulk meter shall apply.


C. S. Phillips,
Chief (W&M),
Standards Branch.


R. W. MacLean,
Director,
Standards Branch.

Reference: A-594

TOKHEIM
MODEL 635TP METER.

