



TRADE AND COMMERCE
CANADA

STANDARDS BRANCH

OTTAWA June 6, 1961.....

APPROVAL LISTING

TOKHEIM - BULK METER

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein have been listed as approved devices and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Volumetric 1 1/2" bulk meter and ancillary equipment, Models 682, 682TP, 683, 684, 684TP, 685 and 685TP, manufactured by Tokheim Corporation, Fort Wayne, Indiana, U.S.A.


Rating of Apparatus: 50 G.P.M. maximum.


Application: Tank truck meter, general trade use and metering liquid fertilizers.

Conditions: As prescribed in P.C. 6894.


Description: These devices are of the same design as meters listed on circular S-WA.403, except that the delivery rate has been increased and larger air eliminator supplied. Model numbers vary due to additional equipment supplied such as: built in filter and air eliminator, ticket printer, non-printing register, built in back pressure check valve and when used for metering fertilizer liquid the letters L.F. are added to the model number as parts of the meter in contact with fertilizer are manufactured from non-corrosive metal.

Testing: The standard tests applicable to a bulk meter shall apply.


C. S. Phillips,
Chief (W&M),
Standards Branch.


R. W. MacLean,
Director,
Standards Branch.

Reference: A-594



(

(