



Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL - AVIS D'APPROBATION

NO.	S.WA-635
	Supplément 1
DATE	February 5, 1973

APPROVED: NEPTUNE - METERS FOR LPG - TRUCK USE

manufactured by: Neptune Meters Limited, Toronto

APPARATUS LISTED: Meters for industrial grades of liquid propane, models as listed below:

<u>MODEL</u>	<u>SIZE</u>	<u>RATING GPM</u>	<u>REMARKS</u>
Type 4D	1 1/4"	5-25	Approved with ATC
Type 4D	1 1/2"	10-50	Approved with ATC
Type 4D	2"	17-85	Approved with ATC; not approved for delivery of under 40 gallons


REVISION: These meters were originally approved when equipped with Neptune Model 431, 432, 433 or 434 registers. They are now also approved when equipped with Veeder-Root registers as follows: large numeral wheel counter; large numeral wheel counter with printer; large numeral wheel counter and set stop; large numeral wheel counter, printer and set stop.


Meter calibration will be accomplished by means of a gear shifter calibrator between the meter and the register, or where vertical clearances do not permit the use of this calibrator, by gear changes in the Veeder-Root register gear plate.

TESTING: The standard tests for a meter for LPG shall apply.

REFERENCES: G 1151-57/N2-30; letter from Neptune of January 29, 1973

CONDITIONS OF APPROVAL: Approval is granted under the Weights and Measures Act, R.S.C., 1970, c. W-7, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above and in the original issue of S.WA-635.

(For) 
Chief
Weights and Measures Division
Standards Branch


R.W. MacLean
Director
Standards Branch