



RECORD OF APPROVAL ACTION

Approval Listing Number-

S.WA-851

Addendum 1

Ottawa, August 27, 1980


Company Requesting Approval: Brooks Instrument Division  
Emerson Electric Canada Limited  
Markham, Ontario

Type of Device: The following meter model is to be added to the  
listing on S.WA-851, dated April 19, 1978.

<u>MODEL NO.</u>	<u>FLOW RATING</u>		<u>MAX. PRESSURE</u>	<u>CONNECTIONS</u>
	<u>G.P.M.</u>	<u>L/M</u>		
B1-43	16-80	75-375	350 p.s.i.	2" Flanges

This meter is of single case construction.

File No.: G6953-B365

.....  
  
 John Armstrong  
 Chief,  
 Weights and Measures Division  
 Legal Metrology Branch