



Consumer and
Corporate Affairs

Consommation et
corporations

Standards

Normes

**NOTICE OF APPROVAL
AVIS D'APPROBATION**

S.WA-908

Ottawa, September 26, 1978

HENDERSON - QUART CAPACITY OIL DISPENSER

MANUFACTURER: Henderson's Minit Dispensers Ltd.
221 Higgins Avenue
Winnipeg, Manitoba

DEVICE APPROVED: Liquid Dispenser, model LD-VIII, capacity 1 quart

APPLICATION: The measurement in trade of automobile lubricants
and anti-freeze.

DESCRIPTION: This dispenser utilizes a cylindrical measuring chamber
of clear plastic material, with a circumferential indicating line
at rated capacity level and the legend "1 quart" marked adjacent to
the line at two diametrically-opposite locations. The liquid to be
dispensed is admitted to the measuring chamber, either by means of
a pressurized supply system or by gravity. The liquid is discharged
by gravity through a manually-operated drain valve. The measuring
chamber has a sloping bottom to facilitate drainage.

TESTING: The standard tests for static volume measures shall apply.

SPECIAL CONDITIONS: As specified in correspondence with the manu-
facturer; copies of this correspondence are on file in the Winnipeg
District Office of Weights and Measures.

REFERENCES: G6953-H192
Test Report, March 6, 1975
Letter of Approval, March 17, 1975

CONDITIONS OF APPROVAL: Approval is granted under the Weights and
Measures Act, S.C. 1970-71-72, Chapter 36, and the Weights and
Measures Regulations P.C. 1974-1461 of June 27, 1974 for use
in Canada under the general conditions of the said Regulations,
and under any special conditions listed above.

John Armstrong
Chief, Weights and Measures
Legal Metrology Branch

HENDERSON - QUART CAPACITY OIL DISPENSER

